



Certificate of Analysis

Material name:	Anchor Covid Influenza RSV Reference Panel
Part number:	60250003
Fill Volume:	1 ml
Lot Number:	401059
Source:	Recombinant RNA Virus harboring SARS-COV-2, Influenza A/B & RSV AB sequence
Storage:	≤ -20 °C Expiration date: 5/31/2026


Testing Results

	Observed Titer (copy/ml)	Observed SD (copy/ml)	Passing Criteria	Result
COVID (N-gene)	9.5×10^5	1.2×10^5	$2.5 \times 10^5 - 1.0 \times 10^6$	PASS
Influenza (A/B)	9.9×10^5	1.6×10^5	$2.5 \times 10^5 - 1.0 \times 10^6$	PASS
RSV A/B	6.9×10^5	5.9×10^4	$2.5 \times 10^5 - 1.0 \times 10^6$	PASS

Approval

Name: Yabin Lu

Signature:

DocuSigned by:

 A68122C2A4CC4F1...

Date: 5/25/2024

Caution

The material is for research use only.